

ICE Checklist #4: Cleaning and Disinfection of the Dialysis Station

Certification Number: _____

Observation 1: Shift # _____ Staff Type _____ Isolation Y ☐ N ☐ Visible From Nursing Station Y ☐ N ☐

Observation 2: Shift # _____ Staff Type _____ Isolation Y ☐ N ☐ Visible From Nursing Station Y ☐ N ☐

Note: In other health care settings, patients vacate treatment area before cleaning and disinfection.

Patient should be vacated from station before cleaning/disinfection of the machine/station unless contraindicated by patient condition. Clinical judgment must be exercised to determine appropriate practice for each patient, ensuring that the patient is fully stabilized prior to discharge.

Was the dialysis station vacated prior to cleaning/disinfection? Obs 1: Y ☐ N ☐ Obs 2: Y ☐ N ☐

Machine: Don gown, gloves, impermeable mask/eye protection or face shield; remove all bloodlines and disposable equipment and discard in biohazardous waste; reprocess dialyzer, with all ports capped; transport dialyzer and bloodlines in a manner that prevents contamination of other surfaces; remove gloves

Obs 1: Met ☐ Not Met ☐

Obs 2: Met ☐ Not Met ☐

Hand hygiene

Obs 1: Met ☐ Not Met ☐

Obs 2: Met ☐ Not Met ☐

Don clean gloves; obtain EPA-registered disinfectant; use tuberculocidal disinfectant if blood is visible

Obs 1: Met ☐ Not Met ☐

Obs 2: Met ☐ Not Met ☐

Wipe all machine top, front, and side surfaces and dialysate hoses wet with disinfectant per manufacturer directions for use; if blood is visible, do second application of tuberculocidal disinfectant per manufacturer directions for use

Obs 1: Met ☐ Not Met ☐

Obs 2: Met ☐ Not Met ☐

Empty prime waste receptacle; wipe all internal and external surfaces wet with disinfectant per manufacturer directions for use

Obs 1: Met ☐ Not Met ☐

Obs 2: Met ☐ Not Met ☐

Chair: Vacated, fully reclined, all disposable supplies removed and discarded; with new disinfectant, wipe all external front-facing and side chair surfaces wet with disinfectant per manufacturer directions for use, including down sides of seat cushion and side tables

Obs 1: Met ☐ Not Met ☐

Obs 2: Met ☐ Not Met ☐

Nondisposable items: Blood pressure cuff, TV controls, call button, data entry station, and counters around station are cleaned and wiped wet with disinfectant

Obs 1: Met ☐ Not Met ☐

Obs 2: Met ☐ Not Met ☐

If clamps are used, clean of visible blood and dirt and disinfect

Obs 1: Met ☐ Not Met ☐

Obs 2: Met ☐ Not Met ☐

Discard cloth/wipe; remove gloves

Obs 1: Met ☐ Not Met ☐

Obs 2: Met ☐ Not Met ☐

Hand hygiene

Obs 1: Met ☐ Not Met ☐

Obs 2: Met ☐ Not Met ☐

Observation 1 notes:

Observation 2 notes: